

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032025\
 Data File : VN086030.D
 Acq On : 20 Mar 2025 17:33
 Operator : JC\MD
 Sample : Q1575-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW7

Quant Time: Mar 21 02:47:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	138616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	242981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	226759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104495	55.052	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.100%
35) Dibromofluoromethane	8.165	113	89735	52.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.820%
50) Toluene-d8	10.565	98	306504	49.796	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.600%
62) 4-Bromofluorobenzene	12.847	95	110086	50.150	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.300%
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.483	96	1966	1.115	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032025\
Data File : VN086030.D
Acq On : 20 Mar 2025 17:33
Operator : JC\MD
Sample : Q1575-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW7

Quant Time: Mar 21 02:47:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

