

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054517.D
 Acq On : 21 Mar 2019 18:00
 Operator : JC/SP
 Sample : K1965-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-HN-MW24IR-20190315

Quant Time: Mar 22 10:45:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	459068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	847476	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	899062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	344453	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	310434	54.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.74%	
35) Dibromofluoromethane	7.59	113	263824	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
50) Toluene-d8	10.09	98	1094081	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
62) 4-Bromofluorobenzene	12.40	95	426866	61.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.24%	
Target Compounds						
7) Trichlorofluoromethane	3.03	101	9203	1.264	ug/l	94
12) 1,1-Dichloroethene	3.73	96	3718	0.840	ug/l	88
19) Methyl tert-butyl Ether	5.06	73	15703	1.215	ug/l	98
24) 1,1-Dichloroethane	5.85	63	5514	0.623	ug/l #	80
44) Trichloroethene	8.83	130	22939	3.641	ug/l	92
64) Tetrachloroethene	10.63	164	18836	2.526	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032119\
Data File : VN054517.D
Acq On : 21 Mar 2019 18:00
Operator : JC/SP
Sample : K1965-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-HN-MW24IR-20190315

Quant Time: Mar 22 10:45:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

