

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054524.D
 Acq On : 21 Mar 2019 21:12
 Operator : JC/SP
 Sample : K1964-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-11

Quant Time: Mar 22 11:06:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	448549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	815220	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	849324	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	300050	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	298478	53.50	ug/l	0.00
Spiked Amount						
			Recovery	=	107.00%	
35) Dibromofluoromethane	7.59	113	258203	49.16	ug/l	0.00
Spiked Amount						
			Recovery	=	98.32%	
50) Toluene-d8	10.09	98	1068106	54.10	ug/l	0.00
Spiked Amount						
			Recovery	=	108.20%	
62) 4-Bromofluorobenzene	12.40	95	390450	58.11	ug/l	0.00
Spiked Amount						
			Recovery	=	116.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032119\
Data File : VN054524.D
Acq On : 21 Mar 2019 21:12
Operator : JC/SP
Sample : K1964-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-11

Quant Time: Mar 22 11:06:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

