

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054533.D
 Acq On : 22 Mar 2019 2:15
 Operator : JC/SP
 Sample : K1688-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-031919-A

Quant Time: Mar 22 10:42:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	663184	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1023503	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	831072	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	299856	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	414999	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
35) Dibromofluoromethane	7.59	113	324279	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
50) Toluene-d8	10.09	98	1194825	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
62) 4-Bromofluorobenzene	12.40	95	361409	42.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.68%	
Target Compounds						
16) Acetone	3.82	43	23194	9.352	ug/l	88
20) Methylene Chloride	4.55	84	251603	34.416	ug/l #	86
30) Chloroform	7.37	83	7150	0.573	ug/l	85
43) Isopropyl Acetate	8.18	43	18782	1.634	ug/l #	91
95) Naphthalene	15.13	128	13435	1.256	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032119\
Data File : VN054533.D
Acq On : 22 Mar 2019 2:15
Operator : JC/SP
Sample : K1688-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-031919-A

Quant Time: Mar 22 10:42:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

