

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054535.D
 Acq On : 22 Mar 2019 3:11
 Operator : JC/SP
 Sample : K1694-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-032019-C

Quant Time: Mar 22 10:48:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	640199	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1015067	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	838390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	310834	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	407079	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
35) Dibromofluoromethane	7.59	113	316653	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
50) Toluene-d8	10.09	98	1194327	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
62) 4-Bromofluorobenzene	12.40	95	369186	44.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.26%	
Target Compounds						
20) Methylene Chloride	4.55	84	18649	2.643	ug/l #	81
30) Chloroform	7.38	83	7802	0.647	ug/l #	81
43) Isopropyl Acetate	8.17	43	19209	1.685	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032119\
Data File : VN054535.D
Acq On : 22 Mar 2019 3:11
Operator : JC/SP
Sample : K1694-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-02-032019-C

Quant Time: Mar 22 10:48:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

