

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054537.D
 Acq On : 22 Mar 2019 4:06
 Operator : JC/SP
 Sample : K1694-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-02-032019-A

Quant Time: Mar 22 10:53:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	645335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1015970	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	830295	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	306622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	407170	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
35) Dibromofluoromethane	7.59	113	312373	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
50) Toluene-d8	10.09	98	1202025	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
62) 4-Bromofluorobenzene	12.40	95	366050	43.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.44%	
Target Compounds						
16) Acetone	3.83	43	14919	6.182	ug/l #	81
20) Methylene Chloride	4.54	84	3970357	558.110	ug/l #	84
43) Isopropyl Acetate	8.17	43	16408	1.438	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	33631	1.693	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032119\
Data File : VN054537.D
Acq On : 22 Mar 2019 4:06
Operator : JC/SP
Sample : K1694-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-02-032019-A

Quant Time: Mar 22 10:53:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

