

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054538.D
 Acq On : 22 Mar 2019 4:33
 Operator : JC/SP
 Sample : K2010-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Mar 22 10:54:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	664023	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1048059	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	862033	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	311900	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	420528	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
35) Dibromofluoromethane	7.59	113	326131	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
50) Toluene-d8	10.09	98	1228442	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
62) 4-Bromofluorobenzene	12.40	95	377994	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
Target Compounds						
16) Acetone	3.83	43	15340	6.178	ug/l #	83
20) Methylene Chloride	4.55	84	51323	7.011	ug/l #	82
43) Isopropyl Acetate	8.18	43	16774	1.425	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032119\
Data File : VN054538.D
Acq On : 22 Mar 2019 4:33
Operator : JC/SP
Sample : K2010-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Mar 22 10:54:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

