

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054539.D
 Acq On : 22 Mar 2019 5:01
 Operator : JC/SP
 Sample : K1994-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 50015-51054-30199-51048-51050

Quant Time: Mar 22 10:56:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	621905	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	989854	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	818926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	302626	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	397887	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
35) Dibromofluoromethane	7.59	113	305126	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
50) Toluene-d8	10.09	98	1172816	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	12.40	95	356547	43.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.42%	
Target Compounds						
20) Methylene Chloride	4.54	84	133333	19.449	ug/l #	85
43) Isopropyl Acetate	8.17	43	16806	1.512	ug/l #	94
95) Naphthalene	15.13	128	50138	4.645	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032119\
Data File : VN054539.D
Acq On : 22 Mar 2019 5:01
Operator : JC/SP
Sample : K1994-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
50015-51054-30199-51048-51050

Quant Time: Mar 22 10:56:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

