

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN032119\
 Data File : VN054528.D
 Acq On : 21 Mar 2019 23:30
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 10:31:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.478	0.507	-6.1	97	0.00
3 P	Chloromethane	0.650	0.681	-4.8	97	0.00
4 C	Vinyl Chloride	0.653	0.671	-2.8#	94	0.00
5 T	Bromomethane	0.422	0.369	12.6	87	0.00
6 T	Chloroethane	0.392	0.379	3.3	92	0.00
7 T	Trichlorofluoromethane	0.793	0.817	-3.0	95	0.00
8 T	Diethyl Ether	0.305	0.297	2.6	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.466	0.428	8.2	88	0.00
10 T	Methyl Iodide	0.686	0.611	10.9	80	0.00
11 T	Tert butyl alcohol	0.038	0.045	-18.4	105	0.00
12 CM	1,1-Dichloroethene	0.482	0.457	5.2#	89	0.00
13 T	Acrolein	0.052	0.057	-9.6	99	0.00
14 T	Allyl chloride	0.908	0.999	-10.0	102	0.00
15 T	Acrylonitrile	0.188	0.225	-19.7	107	0.00
16 T	Acetone	0.187	0.166	11.2	78	0.00
17 T	Carbon Disulfide	1.292	1.284	0.6	96	0.00
18 T	Methyl Acetate	0.417	0.502	-20.4	110	0.00
19 T	Methyl tert-butyl Ether	1.407	1.590	-13.0	103	0.00
20 T	Methylene Chloride	0.551	0.575	-4.4	101	0.00
21 T	trans-1,2-Dichloroethene	0.492	0.546	-11.0	101	0.00
22 T	Diisopropyl ether	1.791	2.057	-14.9	105	0.00
23 T	Vinyl Acetate	1.265	1.529	-20.9	104	0.00
24 P	1,1-Dichloroethane	0.964	1.084	-12.4	103	0.00
25 T	2-Butanone	0.240	0.287	-19.6	102	0.00
26 T	2,2-Dichloropropane	0.688	0.647	6.0	85	0.00
27 T	cis-1,2-Dichloroethene	0.587	0.636	-8.3	101	0.00
28 T	Bromochloromethane	0.475	0.563	-18.5	108	0.00
29 T	Tetrahydrofuran	0.155	0.199	-28.4#	109	0.00
30 C	Chloroform	0.941	1.050	-11.6#	102	0.00
31 T	Cyclohexane	1.121	1.003	10.5	101	0.00
32 T	1,1,1-Trichloroethane	0.777	0.877	-12.9	101	0.00
33 S	1,2-Dichloroethane-d4	0.622	0.703	-13.0	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.322	0.335	-4.0	99	0.00
36 T	1,1-Dichloropropene	0.468	0.466	0.4	100	0.00
37 T	Ethyl Acetate	0.333	0.381	-14.4	108	0.00
38 T	Carbon Tetrachloride	0.454	0.450	0.9	99	0.00
39 T	Methylcyclohexane	0.535	0.500	6.5	94	0.00
40 TM	Benzene	1.337	1.377	-3.0	101	0.00
41 T	Methacrylonitrile	0.203	0.226	-11.3	111	0.00
42 TM	1,2-Dichloroethane	0.487	0.501	-2.9	103	0.00
43 T	Isopropyl Acetate	0.562	0.638	-13.5	106	0.00
44 TM	Trichloroethene	0.372	0.367	1.3	99	0.00
45 C	1,2-Dichloropropane	0.359	0.388	-8.1#	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN032119\
 Data File : VN054528.D
 Acq On : 21 Mar 2019 23:30
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 10:31:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.226	-5.6	102	0.00
47 T	Bromodichloromethane	0.444	0.473	-6.5	103	0.00
48 T	Methyl methacrylate	0.305	0.334	-9.5	106	0.00
49 T	1,4-Dioxane	0.003	0.004	-33.3#	102	0.00
50 S	Toluene-d8	1.211	1.226	-1.2	98	0.00
51 T	4-Methyl-2-Pentanone	0.311	0.362	-16.4	107	0.00
52 CM	Toluene	0.822	0.819	0.4#	98	0.00
53 T	t-1,3-Dichloropropene	0.414	0.447	-8.0	98	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.541	-6.1	99	0.00
55 T	1,1,2-Trichloroethane	0.296	0.310	-4.7	101	0.00
56 T	Ethyl methacrylate	0.397	0.438	-10.3	102	0.00
57 T	1,3-Dichloropropane	0.514	0.546	-6.2	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.163	0.172	-5.5	97	0.00
59 T	2-Hexanone	0.203	0.230	-13.3	105	0.00
60 T	Dibromochloromethane	0.315	0.335	-6.3	99	0.00
61 T	1,2-Dibromoethane	0.292	0.308	-5.5	100	0.00
62 S	4-Bromofluorobenzene	0.412	0.411	0.2	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.415	0.404	2.7	96	0.00
65 PM	Chlorobenzene	1.017	1.013	0.4	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.393	-7.7	101	0.00
67 C	Ethyl Benzene	1.796	1.794	0.1#	97	0.00
68 T	m/p-Xylenes	0.680	0.669	1.6	95	0.00
69 T	o-Xylene	0.664	0.661	0.5	96	0.00
70 T	Styrene	1.079	1.101	-2.0	96	0.00
71 P	Bromoform	0.222	0.247	-11.3	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.029	3.980	1.2	95	0.00
74 T	N-amyl acetate	1.159	1.380	-19.1	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.920	1.037	-12.7	102	0.00
76 T	1,2,3-Trichloropropane	0.793	0.949	-19.7	109	0.00
77 T	Bromobenzene	0.999	1.014	-1.5	96	0.00
78 T	n-propylbenzene	4.369	4.354	0.3	94	0.00
79 T	2-Chlorotoluene	2.724	2.697	1.0	96	0.00
80 T	1,3,5-Trimethylbenzene	3.302	3.291	0.3	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.205	0.228	-11.2	96	0.00
82 T	4-Chlorotoluene	2.641	2.663	-0.8	94	0.00
83 T	tert-Butylbenzene	2.874	2.875	-0.0	95	0.00
84 T	1,2,4-Trimethylbenzene	3.240	3.263	-0.7	93	0.00
85 T	sec-Butylbenzene	3.583	3.535	1.3	94	0.00
86 T	p-Isopropyltoluene	3.206	3.173	1.0	92	0.00
87 T	1,3-Dichlorobenzene	1.656	1.653	0.2	91	0.00
88 T	1,4-Dichlorobenzene	1.611	1.615	-0.2	92	0.00
89 T	n-Butylbenzene	2.507	2.532	-1.0	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN032119\
Data File : VN054528.D
Acq On : 21 Mar 2019 23:30
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 22 10:31:08 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.498	0.519	-4.2	94	0.00
91 T	1,2-Dichlorobenzene	1.604	1.645	-2.6	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.135	0.157	-16.3	95	0.00
93 T	1,2,4-Trichlorobenzene	0.800	0.835	-4.4	88	0.00
94 T	Hexachlorobutadiene	0.523	0.450	14.0	85	0.00
95 T	Naphthalene	1.783	1.970	-10.5	92	0.00
96 T	1,2,3-Trichlorobenzene	0.796	0.844	-6.0	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6