

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054505.D
 Acq On : 21 Mar 2019 11:57
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

MMDadoda
 3/22/2019 4:15:17 PM

Quant Time: Mar 22 10:16:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	568655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	861072	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	749355	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	341063	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	343180	48.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.04%	
35) Dibromofluoromethane	7.59	113	280671	50.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.18%	
50) Toluene-d8	10.09	98	1058348	50.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.50%	
62) 4-Bromofluorobenzene	12.40	95	366376	51.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.26%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.85	85	280334	51.568	ug/l	99
3) Chloromethane	2.06	50	355998	48.131	ug/l	99
4) Vinyl Chloride	2.19	62	351309	47.308	ug/l	100
5) Bromomethane	2.57	94	203977	42.479	ug/l	96
6) Chloroethane	2.71	64	196929	44.213	ug/l	99
7) Trichlorofluoromethane	3.02	101	427273	47.379	ug/l	99
8) Diethyl Ether	3.41	74	155295	44.775	ug/l	75
9) 1,1,2-Trichlorotrifluoroet	3.75	101	244879	46.203	ug/l	95
10) Methyl Iodide	3.94	142	346137	44.397	ug/l	93
11) Tert butyl alcohol	4.82	59	103757	242.295	ug/l	98
12) 1,1-Dichloroethene	3.73	96	233973	42.658	ug/l #	79
13) Acrolein	3.61	56	133043	225.132	ug/l	98
14) Allyl chloride	4.32	41	511706	49.544	ug/l #	91
15) Acrylonitrile	4.99	53	540476	253.445	ug/l	99
16) Acetone	3.82	43	529988	249.224	ug/l #	85
17) Carbon Disulfide	4.04	76	700541	47.668	ug/l	99
18) Methyl Acetate	4.33	43	230992	48.734	ug/l	91
19) Methyl tert-butyl Ether	5.05	73	779147	48.675	ug/l	94
20) Methylene Chloride	4.55	84	293916	46.887	ug/l #	84
21) trans-1,2-Dichloroethene	5.04	96	279503	49.905	ug/l #	84
22) Diisopropyl ether	5.96	45	1026179	50.371	ug/l #	94
23) Vinyl Acetate	5.90	43	3780346	262.734	ug/l #	93
24) 1,1-Dichloroethane	5.85	63	545610	49.740	ug/l	98
25) 2-Butanone	6.84	43	721837	264.981	ug/l	90
26) 2,2-Dichloropropane	6.82	77	406014	51.924	ug/l	98
27) cis-1,2-Dichloroethene	6.83	96	319336	47.802	ug/l	88
28) Bromochloromethane	7.19	49	260365	48.222	ug/l #	75
29) Tetrahydrofuran	7.22	42	453588	256.588	ug/l #	85
30) Chloroform	7.37	83	521798	48.744	ug/l	100
31) Cyclohexane	7.65	56	543430	51.768	ug/l	86
32) 1,1,1-Trichloroethane	7.57	97	444539	50.335	ug/l	96
36) 1,1-Dichloropropene	7.79	75	417720	51.884	ug/l	96
37) Ethyl Acetate	6.94	43	299739	52.345	ug/l	96
38) Carbon Tetrachloride	7.77	117	395095	50.499	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054505.D
 Acq On : 21 Mar 2019 11:57
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

MMDadoda
 3/22/2019 4:15:17 PM

Quant Time: Mar 22 10:16:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

39) Methylcyclohexane	9.08	83	498994	54.182	ug/l	90
40) Benzene	8.04	78	1178975	51.210	ug/l	99
41) Methacrylonitrile	7.17	41	176792	50.625	ug/l #	89
42) 1,2-Dichloroethane	8.12	62	418074	49.864	ug/l	98
43) Isopropyl Acetate	8.17	43	520945	53.868	ug/l #	94
44) Trichloroethene	8.83	130	320153	50.018	ug/l	93
45) 1,2-Dichloropropane	9.12	63	325697	52.684	ug/l	99
46) Dibromomethane	9.21	93	191496	51.923	ug/l	89
47) Bromodichloromethane	9.40	83	402962	52.647	ug/l	97
48) Methyl methacrylate	9.20	41	266524	50.777	ug/l	88
49) 1,4-Dioxane	9.20	88	58720	1000.682	ug/l #	84
51) 4-Methyl-2-Pentanone	9.99	43	1454816	271.442	ug/l	94
52) Toluene	10.16	92	723043	51.062	ug/l	99
53) t-1,3-Dichloropropene	10.38	75	399106	49.456	ug/l	98
54) cis-1,3-Dichloropropene	9.84	75	480040	54.698	ug/l	92
55) 1,1,2-Trichloroethane	10.56	97	261013	51.280	ug/l	97
56) Ethyl methacrylate	10.43	69	371457	54.372	ug/l #	86
57) 1,3-Dichloropropane	10.71	76	462674	52.291	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.70	63	717013	254.877	ug/l	91
59) 2-Hexanone	10.75	43	958580	273.817	ug/l	94
60) Dibromochloromethane	10.90	129	291187	53.762	ug/l	99
61) 1,2-Dibromoethane	11.00	107	260932	51.854	ug/l	99
64) Tetrachloroethene	10.63	164	300227	48.302	ug/l	96
65) Chlorobenzene	11.43	112	770625	50.563	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.51	131	286713	52.411	ug/l	99
67) Ethyl Benzene	11.51	91	1368336	50.826	ug/l	99
68) m/p-Xylenes	11.62	106	1043449	102.324	ug/l	97
69) o-Xylene	11.95	106	498659	50.072	ug/l	97
70) Styrene	11.96	104	837597	51.791	ug/l	98
71) Bromoform	12.13	173	179650	47.739	ug/l #	99
73) Isopropylbenzene	12.25	105	1335371	48.588	ug/l	98
74) N-amyl acetate	12.07	43	436124	55.188	ug/l	93
75) 1,1,2,2-Tetrachloroethane	12.50	83	319067	50.857	ug/l	99
76) 1,2,3-Trichloropropane	12.55	75	305297m	56.410	ug/l	
77) Bromobenzene	12.53	156	328532	48.198	ug/l	87
78) n-propylbenzene	12.59	91	1508813	50.628	ug/l	97
79) 2-Chlorotoluene	12.67	91	894271	48.119	ug/l	96
80) 1,3,5-Trimethylbenzene	12.73	105	1124383	49.917	ug/l	98
81) trans-1,4-Dichloro-2-buten	12.30	75	75071	47.956	ug/l	86
82) 4-Chlorotoluene	12.77	91	892648	49.553	ug/l	97
83) tert-Butylbenzene	12.99	119	972255	49.587	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	1102362	49.876	ug/l	99
85) sec-Butylbenzene	13.17	105	1241193	50.781	ug/l	97
86) p-Isopropyltoluene	13.29	119	1135248	51.916	ug/l	98
87) 1,3-Dichlorobenzene	13.28	146	559982	49.575	ug/l	98
88) 1,4-Dichlorobenzene	13.36	146	544726	49.564	ug/l	98
89) n-Butylbenzene	13.62	91	950376	55.583	ug/l	99
90) Hexachloroethane	13.87	117	183816	54.155	ug/l	89
91) 1,2-Dichlorobenzene	13.65	146	544119	49.732	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	14.27	75	48100	52.270	ug/l	85

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032119\
Data File : VN054505.D
Acq On : 21 Mar 2019 11:57
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM

Quant Time: Mar 22 10:16:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.91	180	305476	56.008	ug/l	99
94) Hexachlorobutadiene	15.01	225	187717	61.845	ug/l	98
95) Naphthalene	15.13	128	656665	53.977	ug/l	99
96) 1,2,3-Trichlorobenzene	15.31	180	295838	54.499	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032119\
 Data File : VN054505.D
 Acq On : 21 Mar 2019 11:57
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

MMDadoda
 3/22/2019 4:15:17 PM

Quant Time: Mar 22 10:16:01 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion:168 Resp: 568655

Ion Ratio Lower Upper

168 100


99 53.8 37.1 55.7

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.66

250000

200000



150000

100000

50000

0

Time--> 7.50 7.60 7.70 7.80



#2

Dichlorodifluoromethane

Concen: 51.568 ug/l

RT: 1.85 min Scan# 25

Delta R.T. 0.00 min

Lab File: VN054505.D


Acq: 21 Mar 2019 11:57

Tgt Ion: 85 Resp: 280334

Ion Ratio Lower Upper

85 100

87 32.5 15.9 47.7



Abundance Ion 84.90 (84.60 to 85.60): VNO

Ion 86.90 (86.60 to 87.60): VNO

1.85

200000


150000


100000

50000

0

Time--> 1.80 1.85 1.90 1.95





#3

Chloromethane

Concen: 48.131 ug/l

RT: 2.06 min Scan# 90

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 50 Resp: 355998

Ion Ratio Lower Upper

50 100


52 32.6 26.4 39.6

Manual Integrations

APPROVED




MMDadoda

3/22/2019 4:15:17 PM



Abundance Ion 49.90 (49.60 to 50.60): VNO

Ion 51.90 (51.60 to 52.60): VNO



#4

Vinyl Chloride

Concen: 47.308 ug/l

RT: 2.19 min Scan# 130

Delta R.T. 0.00 min

Lab File: VN054505.D


Acq: 21 Mar 2019 11:57

Tgt Ion: 62 Resp: 351309

Ion Ratio Lower Upper



62 100


64 31.7 25.5 38.3



Abundance Ion 61.90 (61.60 to 62.60): VNO

Ion 63.90 (63.60 to 64.60): VNO





#5

Bromomethane

Concen: 42.479 ug/l

RT: 2.57 min Scan# 248

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 94 Resp: 203977

Ion Ratio Lower Upper

94 100


96 92.4 77.0 115.4

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



Abundance Ion 93.90 (93.60 to 94.60): VN0

Ion 95.90 (95.60 to 96.60): VN0

2.57

100000

80000



60000

40000

20000

0

Time-->



#6

Chloroethane

Concen: 44.213 ug/l

RT: 2.71 min Scan# 291

Delta R.T. 0.00 min

Lab File: VN054505.D


Acq: 21 Mar 2019 11:57

Tgt Ion: 64 Resp: 196929

Ion Ratio Lower Upper

64 100

66 32.9 25.7 38.5



Abundance Ion 63.90 (63.60 to 64.60): VN0

Ion 65.90 (65.60 to 66.60): VN0

2.71

100000

80000


60000


40000

20000

0

Time-->





#7





Trichlorofluoromethane
 Concen: 47.379 ug/l
 RT: 3.02 min Scan# 389
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
101	100		
103	63.1	51.4	77.2

Manual Integrations
 APPROVED




MMDadoda
 3/22/2019 4:15:17 PM




#8

Diethyl Ether
 Concen: 44.775 ug/l
 RT: 3.41 min Scan# 510
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
74	100		
45	109.0	43.1	129.4





#9

1,1,2-Trichlorotrifluoroethane

Concen: 46.203 ug/l

RT: 3.75 min Scan# 615

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion:101 Resp: 244879

Ion Ratio Lower Upper

101 100

85 44.2 32.4 48.6


151 81.5 68.2 102.2

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM






Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): V

Ion 150.80 (150.50 to 151.50): V

Time-->



#10

Methyl Iodide

Concen: 44.397 ug/l

RT: 3.94 min Scan# 675

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57


Tgt Ion:142 Resp: 346137

Ion Ratio Lower Upper

142 100

127 43.1 30.3 45.5

141 14.4 11.2 16.8






Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V

Time-->









#11
Tert butyl alcohol
Concen: 242.295 ug/l
RT: 4.82 min Scan# 947
Delta R.T. -0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion: 59 Resp: 103757
Ion Ratio Lower Upper
59 100
57 10.4 9.0 13.6




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#12
1,1-Dichloroethene
Concen: 42.658 ug/l
RT: 3.73 min Scan# 609
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion: 96 Resp: 233973
Ion Ratio Lower Upper
96 100
61 183.4 116.5 174.7#
98 63.4 51.0 76.4





#13

Acrolein

Concen: 225.132 ug/l

RT: 3.61 min Scan# 570

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 56 Resp: 133043

Ion Ratio Lower Upper

56 100





55 72.3 56.7 85.1

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#14

Allyl chloride

Concen: 49.544 ug/l

RT: 4.32 min Scan# 791

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57




Tgt Ion: 41 Resp: 511706


Ion Ratio Lower Upper

41 100

39 67.6 58.2 87.4

76 30.2 31.7 47.5#









#15
 Acrylonitrile
 Concen: 253.445 ug/l
 RT: 4.99 min Scan# 1001
 Delta R.T. -0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
53	100		
52	81.9	66.4	99.6
51	37.0	29.6	44.4




Manual Integrations
 APPROVED


MMDadoda
 3/22/2019 4:15:17 PM



#16
 Acetone
 Concen: 249.224 ug/l
 RT: 3.82 min Scan# 637
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
43	100		
58	28.5	29.9	44.9#









#17
Carbon Disulfide
Concen: 47.668 ug/l
RT: 4.04 min Scan# 705
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
76	100		
78	9.2	7.5	11.3




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#18
Methyl Acetate
Concen: 48.734 ug/l
RT: 4.33 min Scan# 795
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
43	100		
74	21.4	20.8	31.2









#19
Methyl tert-butyl Ether
Concen: 48.675 ug/l
RT: 5.05 min Scan# 1020
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion: 73 Resp: 779147
Ion Ratio Lower Upper
73 100
57 23.9 16.8 25.2




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#20
Methylene Chloride
Concen: 46.887 ug/l
RT: 4.55 min Scan# 862
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion: 84 Resp: 293916
Ion Ratio Lower Upper
84 100
49 138.1 89.7 134.5#
51 42.7 28.2 42.4#
86 64.8 50.8 76.2





#21

trans-1,2-Dichloroethene

Concen: 49.905 ug/l

RT: 5.04 min Scan# 1015

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 96 Resp: 279503

Ion Ratio Lower Upper

96 100

61 154.3 102.8 154.2#





98 64.4 50.0 75.0

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#22

Diisopropyl ether

Concen: 50.371 ug/l

RT: 5.96 min Scan# 1302

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Tgt Ion: 45 Resp: 1026179




Ion Ratio Lower Upper


45 100

43 58.2 48.2 72.2

87 24.1 24.9 37.3#

59 10.7 9.5 14.3









#23
 Vinyl Acetate
 Concen: 262.734 ug/l
 RT: 5.90 min Scan# 1283
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tot Ion: 43 Resp: 3780346
 Ion Ratio Lower Upper
 43 100
 86 8.8 9.1 13.7#




Manual Integrations
 APPROVED


MMDadoda
 3/22/2019 4:15:17 PM



#24
 1,1-Dichloroethane
 Concen: 49.740 ug/l
 RT: 5.85 min Scan# 1267
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Tgt Ion: 63 Resp: 545610
 Ion Ratio Lower Upper
 63 100
 98 6.0 3.5 10.5
 100 3.8 2.0 6.0





#25

2-Butanone

Concen: 264.981 ug/l

RT: 6.84 min Scan# 1577

Delta R.T. -0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 43 Resp: 721837

Ion Ratio Lower Upper

43 100





72 23.5 23.0 34.6

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#26

2,2-Dichloropropane

Concen: 51.924 ug/l

RT: 6.82 min Scan# 1571

Delta R.T. 0.00 min

Lab File: VN054505.D




Acq: 21 Mar 2019 11:57


Tgt Ion: 77 Resp: 406014

Ion Ratio Lower Upper

77 100

97 22.9 11.9 35.9





#27

cis-1,2-Dichloroethene

Concen: 47.802 ug/l

RT: 6.83 min Scan# 1573

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N





ClientSampleId :

VSTDCCC050

Tot Ion	96	Resp	319336
Ion	Ratio	Lower	Upper
96	100		
61	158.2	0.0	273.0
98	65.0	0.0	130.6

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM



#28

Bromochloromethane

Concen: 48.222 ug/l




RT: 7.19 min Scan# 1686


Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Tgt Ion	49	Resp	260365
Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	4.6
130	71.4	76.7	115.1#









#29
Tetrahydrofuran
Concen: 256.588 ug/l
RT: 7.22 min Scan# 1694
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
42	100		
72	37.4	37.8	56.6#
71	35.3	36.2	54.2#




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#30
Chloroform
Concen: 48.744 ug/l
RT: 7.37 min Scan# 1741
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
83	100		
85	65.1	52.2	78.4









#31
Cyclohexane
Concen: 51.768 ug/l
RT: 7.65 min Scan# 1828
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
56	100		
69	30.0	27.8	41.6
84	77.9	74.6	111.8




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#32
1,1,1-Trichloroethane
Concen: 50.335 ug/l
RT: 7.57 min Scan# 1802
Delta R.T. -0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
97	100		
99	63.7	51.5	77.3
61	48.0	33.6	50.4









#33
 1,2-Dichloroethane-d4
 Concen: 48.519 ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
65	100	343180		
67	53.6	0.0	104.6	




Manual Integrations
 APPROVED


MMDadoda
 3/22/2019 4:15:17 PM



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Resp	Lower	Upper
114	100	861072		
63	21.0	0.0	35.2	
88	15.6	0.0	30.2	









#35
 Dibromofluoromethane
 Concen: 50.595 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Resp	Lower	Upper
113	100			
111	101.3	82.2	123.4	
192	21.4	19.2	28.8	




Manual Integrations
 APPROVED


MMDadoda
 3/22/2019 4:15:17 PM



#36
 1,1-Dichloropropene
 Concen: 51.884 ug/l
 RT: 7.79 min Scan# 1871
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Resp	Lower	Upper
75	100			
110	34.0	19.1	57.5	
77	30.5	25.0	37.4	









#37
Ethyl Acetate
Concen: 52.345 ug/l
RT: 6.94 min Scan# 1606
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
43	100		
61	13.4	11.8	17.6
70	9.1	8.4	12.6




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#38
Carbon Tetrachloride
Concen: 50.499 ug/l
RT: 7.77 min Scan# 1865
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
117	100		
119	94.0	76.3	114.5
121	30.8	24.6	36.8









#39
Methvlcvclohexane
Concen: 54.182 ug/l
RT: 9.08 min Scan# 2272
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
83	100		
55	82.4	56.6	84.8
98	44.9	38.4	57.6




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#40
Benzene
Concen: 51.210 ug/l
RT: 8.04 min Scan# 1949
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
78	100		
77	23.5	19.1	28.7






#41
Methacrylonitrile
Concen: 50.625 ug/l
RT: 7.17 min Scan# 1679
Delta R.T. -0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
41	100		
39	57.2	46.8	70.2
67	66.9	67.0	100.4
52	29.0	26.5	39.7

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM






Abundance

Ion 41.00 (40.70 to 41.70): VNO

Ion 39.00 (38.70 to 39.70): VNO


Ion 67.00 (66.70 to 67.70): VNO

Ion 52.00 (51.70 to 52.70): VNO



#42
1,2-Dichloroethane
Concen: 49.864 ug/l
RT: 8.12 min Scan# 1975
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57



Tgt Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	18.4




Abundance

Ion 61.90 (61.60 to 62.60): VNO

Ion 97.90 (97.60 to 98.60): VNO









#43
Isopropyl Acetate
Concen: 53.868 ug/l
RT: 8.17 min Scan# 1989
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
43	100		
61	18.3	15.8	23.8
87	11.4	12.0	18.0




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#44
Trichloroethene
Concen: 50.018 ug/l
RT: 8.83 min Scan# 2196
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
130	100		
95	99.5	0.0	185.2





#45

1,2-Dichloropropane

Concen: 52.684 ug/l

RT: 9.12 min Scan# 2285

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 63 Resp: 325697

Ion Ratio Lower Upper

63 100


65 30.9 25.0 37.6

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 64.90 (64.60 to 65.60): VN0

9.12



150000

100000

50000

0

Time-->



#46

Dibromomethane

Concen: 51.923 ug/l

RT: 9.21 min Scan# 2313

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57


Tgt Ion: 93 Resp: 191496

Ion Ratio Lower Upper

93 100

95 83.1 66.7 100.1

174 103.3 98.5 147.7



Abundance Ion 92.80 (92.50 to 93.50): VN0

Ion 94.80 (94.50 to 95.50): VN0

Ion 173.70 (173.40 to 174.40): V


9.21


100000

50000

0

Time-->





#47

Bromodichloromethane

Concen: 52.647 ug/l

RT: 9.40 min Scan# 2373

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N





ClientSampleId :

VSTDCCC050

Tot Ion:	83	Resp:	402962
Ion Ratio	Lower	Upper	
83	100		
85	63.3	52.8	79.2
127	9.2	7.4	11.2

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM



#48

Methyl methacrylate

Concen: 50.777 ug/l




RT: 9.20 min Scan# 2310


Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Tgt Ion:	41	Resp:	266524
Ion Ratio	Lower	Upper	
41	100		
69	74.8	72.6	108.8
39	55.9	47.5	71.3





#49

1,4-Dioxane

Concen: 1000.682 ug/l

RT: 9.20 min Scan# 2310

Delta R.T. -0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 88 Resp: 58720

Ion Ratio Lower Upper

88 100

43 38.4 24.1 36.1#





58 76.7 50.8 76.2#

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#50

Toluene-d8

Concen: 50.748 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN054505.D




Acq: 21 Mar 2019 11:57


Tgt Ion: 98 Resp: 1058348

Ion Ratio Lower Upper

98 100

100 64.6 51.2 76.8





#51

4-Methyl-2-Pentanone

Concen: 271.442 ug/l

RT: 9.99 min Scan# 2554

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 43 Resp: 1454816

Ion Ratio Lower Upper

43 100





58 36.3 31.8 47.6

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#52

Toluene

Concen: 51.062 ug/l

RT: 10.16 min Scan# 2607

Delta R.T. 0.00 min

Lab File: VN054505.D




Acq: 21 Mar 2019 11:57


Tgt Ion: 92 Resp: 723043

Ion Ratio Lower Upper

92 100

91 174.1 140.7 211.1





#53

t-1,3-Dichloropropene

Concen: 49.456 ug/l

RT: 10.38 min Scan# 2677

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 75 Resp: 399106

Ion Ratio Lower Upper

75 100





77 31.2 25.8 38.8

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#54

cis-1,3-Dichloropropene

Concen: 54.698 ug/l

RT: 9.84 min Scan# 2508

Delta R.T. -0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57




Tgt Ion: 75 Resp: 480040


Ion Ratio Lower Upper

75 100

77 31.3 25.4 38.0

39 54.9 37.2 55.8





#55

1,1,2-Trichloroethane

Concen: 51.280 ug/l

RT: 10.56 min Scan# 2733

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N





ClientSampleId :

VSTDCCC050

Tot Ion:	97	Resp:	261013
Ion Ratio	Lower	Upper	
97	100		
83	88.3	66.6	100.0
85	56.7	44.4	66.6
99	62.6	50.8	76.2

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM



#56

Ethyl methacrylate

Concen: 54.372 ug/l




RT: 10.43 min Scan# 2692


Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Tgt Ion:	69	Resp:	371457
Ion Ratio	Lower	Upper	
69	100		
41	75.3	49.4	74.2#
39	39.9	28.3	42.5





#57

1,3-Dichloropropane

Concen: 52.291 ug/l

RT: 10.71 min Scan# 2779

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 76 Resp: 462674

Ion Ratio Lower Upper

76 100





78 32.0 25.7 38.5

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#58

2-Chloroethyl Vinyl ether

Concen: 254.877 ug/l

RT: 9.70 min Scan# 2464

Delta R.T. 0.00 min

Lab File: VN054505.D




Acq: 21 Mar 2019 11:57


Tgt Ion: 63 Resp: 717013

Ion Ratio Lower Upper

63 100

106 25.8 24.4 36.6









#59
2-Hexanone
Concen: 273.817 ug/l
RT: 10.75 min Scan# 2793
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion: 43 Resp: 958580
Ion Ratio Lower Upper
43 100
58 51.3 28.0 83.9




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#60
Dibromochloromethane
Concen: 53.762 ug/l
RT: 10.90 min Scan# 2839
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion: 129 Resp: 291187
Ion Ratio Lower Upper
129 100
127 76.2 38.6 115.8





#61

1,2-Dibromoethane

Concen: 51.854 ug/l

RT: 11.00 min Scan# 2871

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion:107 Resp: 260932

Ion Ratio Lower Upper

107 100


109 95.5 75.9 113.9

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V

11.00



150000

100000

50000

0

Time--> 10.90 11.00 11.10



#62

4-Bromofluorobenzene

Concen: 51.627 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57


Tgt Ion: 95 Resp: 366376

Ion Ratio Lower Upper

95 100

174 84.6 0.0 191.0

176 82.3 0.0 187.4



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

250000

200000


150000


100000

50000

0

Time--> 12.30 12.40 12.50









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampled :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
117	100		
82	56.6	40.0	60.0
119	32.2	25.8	38.6




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#64
Tetrachloroethene
Concen: 48.302 ug/l
RT: 10.63 min Scan# 2755
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
164	100		
166	129.6	103.9	155.9
129	93.9	71.1	106.7
131	92.2	68.2	102.4









#65
Chlorobenzene
Concen: 50.563 ug/l
RT: 11.43 min Scan# 3004
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion:112 Resp: 770625
Ion Ratio Lower Upper
112 100
114 32.5 25.9 38.9




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#66
1,1,1,2-Tetrachloroethane
Concen: 52.411 ug/l
RT: 11.51 min Scan# 3028
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion:131 Resp: 286713
Ion Ratio Lower Upper
131 100
133 95.9 47.6 142.9
119 65.6 32.4 97.0





#67

Ethyl Benzene

Concen: 50.826 ug/l

RT: 11.51 min Scan# 3028

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 91 Resp: 1368336

Ion Ratio Lower Upper

91 100





106 31.3 25.5 38.3

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#68

m/p-Xylenes

Concen: 102.324 ug/l

RT: 11.62 min Scan# 3062

Delta R.T. -0.00 min

Lab File: VN054505.D




Acq: 21 Mar 2019 11:57


Tgt Ion: 106 Resp: 1043449

Ion Ratio Lower Upper

106 100

91 204.1 159.4 239.2









#69
o-Xylene
Concen: 50.072 ug/l
RT: 11.95 min Scan# 3164
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion:106 Resp: 498659
Ion Ratio Lower Upper
106 100
91 216.7 105.7 317.1




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#70
Styrene
Concen: 51.791 ug/l
RT: 11.96 min Scan# 3169
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion:104 Resp: 837597
Ion Ratio Lower Upper
104 100
78 50.9 38.2 57.2
103 55.1 44.4 66.6






#71
Bromoform
Concen: 47.739 ug/l
RT: 12.13 min Scan# 3220
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tot Ion	Ratio	Lower	Upper
173	100		
175	49.5	24.4	73.2
254	0.1	0.1	0.1#

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM




Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
152	100		
115	57.8	27.2	81.6
150	173.6	0.0	349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V









#73
Isopropylbenzene
Concen: 48.588 ug/l
RT: 12.25 min Scan# 3258
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion:105 Resp: 1335371
Ion Ratio Lower Upper
105 100
120 26.4 13.8 41.3




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#74
N-ethyl acetate
Concen: 55.188 ug/l
RT: 12.07 min Scan# 3202
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion: 43 Resp: 436124
Ion Ratio Lower Upper
43 100
70 38.2 35.4 53.0
55 23.1 19.8 29.6
61 22.1 19.8 29.6





#75

1,1,2,2-Tetrachloroethane

Concen: 50.857 ug/l

RT: 12.50 min Scan# 3337

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampled :

VSTDCCC050

Tot Ion: 83 Resp: 319067
Ion Ratio Lower Upper

83 100


131 10.6 5.8 17.4

85 64.1 32.2 96.6




Manual Integrations
APPROVED

MMDadoda

3/22/2019 4:15:17 PM



Abundance Ion 82.90 (82.60 to 83.60): VN0
Ion 130.80 (130.50 to 131.50): VN0
Ion 84.90 (84.60 to 85.60): VN0



#76

1,2,3-Trichloropropane

Concen: 56.410 ug/l m

RT: 12.55 min Scan# 3353

Delta R.T. 0.00 min


Lab File: VN054505.D

Acq: 21 Mar 2019 11:57



Tgt Ion: 75 Resp: 305297
Ion Ratio Lower Upper


75 100

77 0.0 0.0 0.0



Abundance Ion 74.90 (74.60 to 75.60): VN0
Ion 77.00 (76.70 to 77.70): VN0





#77

Bromobenzene

Concen: 48.198 ug/l

RT: 12.53 min Scan# 3344

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N


ClientSampleId :

VSTDCCC050




Tot Ion:	156	Resp:	328532
Ion Ratio	Lower	Upper	
156	100		
77	189.0	81.5	244.4
158	97.1	49.0	147.2

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM



Abundance Ion 155.80 (155.50 to 156.50): V
Ion 77.00 (76.70 to 77.70): VNO
Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 50.628 ug/l


RT: 12.59 min Scan# 3365

Delta R.T. 0.00 min



Lab File: VN054505.D


Acq: 21 Mar 2019 11:57

Tgt Ion:	91	Resp:	1508813
Ion Ratio	Lower	Upper	
91	100		
120	23.2	12.3	36.9



Abundance Ion 91.00 (90.70 to 91.70): VNO
Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 48.119 ug/l

RT: 12.67 min Scan# 3390

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion: 91 Resp: 894271

Ion Ratio Lower Upper

91 100





126 34.8 18.7 56.1

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#80

1,3,5-Trimethylbenzene

Concen: 49.917 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. 0.00 min

Lab File: VN054505.D




Acq: 21 Mar 2019 11:57


Tgt Ion: 105 Resp: 1124383

Ion Ratio Lower Upper

105 100

120 48.6 25.0 75.0









#81
trans-1,4-Dichloro-2-butene
Concen: 47.956 ug/l
RT: 12.30 min Scan# 3274
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
75	100		
53	113.3	76.2	114.2
89	45.4	39.2	58.8




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#82
4-Chlorotoluene
Concen: 49.553 ug/l
RT: 12.77 min Scan# 3421
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
91	100		
126	34.4	18.1	54.1





#83

tert-Butylbenzene

Concen: 49.587 ug/l

RT: 12.99 min Scan# 3490

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N


ClientSampleId :

VSTDCCC050




Tot Ion:	119	Resp:	972255
Ion Ratio	Lower	Upper	
119	100		
91	63.6	30.9	92.8
134	26.2	13.2	39.6

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM



Abundance Ion 119.00 (118.70 to 119.70): V
Ion 91.00 (90.70 to 91.70): V
Ion 134.00 (133.70 to 134.70): V



#84

1,2,4-Trimethylbenzene

Concen: 49.876 ug/l


RT: 13.04 min Scan# 3504

Delta R.T. 0.00 min



Lab File: VN054505.D


Acq: 21 Mar 2019 11:57

Tgt Ion:	105	Resp:	1102362
Ion Ratio	Lower	Upper	
105	100		
120	46.5	23.4	70.3



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 50.781 ug/l

RT: 13.17 min Scan# 3545

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tot Ion:105 Resp: 1241193

Ion Ratio Lower Upper

105 100





134 20.2 10.7 32.1

Manual Integrations

APPROVED

MMDadoda

3/22/2019 4:15:17 PM



#86

p-Isopropyltoluene

Concen: 51.916 ug/l

RT: 13.29 min Scan# 3581

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57




Tgt Ion:119 Resp: 1135248


Ion Ratio Lower Upper

119 100

134 26.5 13.6 40.8

91 23.3 11.1 33.3





#87

1,3-Dichlorobenzene

Concen: 49.575 ug/l

RT: 13.28 min Scan# 3579

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N





ClientSampleId :

VSTDCCC050

Tot Ion:	146	Resp:	559982
Ion Ratio	Lower	Upper	
146	100		
111	40.2	18.8	56.3
148	64.9	32.2	96.6

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM



#88

1,4-Dichlorobenzene

Concen: 49.564 ug/l




RT: 13.36 min Scan# 3604


Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Tgt Ion:	146	Resp:	544726
Ion Ratio	Lower	Upper	
146	100		
111	38.2	18.1	54.1
148	64.6	32.0	96.0









#89
n-Butylbenzene
Concen: 55.583 ug/l
RT: 13.62 min Scan# 3683
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
91	100		
92	52.6	26.2	78.6
134	26.6	14.3	42.9




Manual Integrations
APPROVED


MMDadoda
3/22/2019 4:15:17 PM



#90
Hexachloroethane
Concen: 54.155 ug/l
RT: 13.87 min Scan# 3763
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
117	100		
201	85.9	48.1	144.4









#91
 1,2-Dichlorobenzene
 Concen: 49.732 ug/l
 RT: 13.65 min Scan# 3695
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
146	100		
111	41.0	19.1	57.3
148	63.8	32.4	97.0




Manual Integrations
 APPROVED


MMDadoda
 3/22/2019 4:15:17 PM



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 52.270 ug/l
 RT: 14.27 min Scan# 3886
 Delta R.T. 0.00 min
 Lab File: VN054505.D
 Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
75	100		
155	86.1	50.1	150.3
157	107.7	63.1	189.3





#93

1,2,4-Trichlorobenzene

Concen: 56.008 ug/l

RT: 14.91 min Scan# 4085

Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Instrument :

MSVOA_N





ClientSampleId :

VSTDCCC050

Tot Ion:	180	Resp:	305476
Ion Ratio	Lower	Upper	
180	100		
182	97.0	48.0	144.2
145	29.6	14.2	42.8

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM



#94

Hexachlorobutadiene

Concen: 61.845 ug/l




RT: 15.01 min Scan# 4117


Delta R.T. 0.00 min

Lab File: VN054505.D

Acq: 21 Mar 2019 11:57

Tgt Ion:	225	Resp:	187717
Ion Ratio	Lower	Upper	
225	100		
223	63.7	31.1	93.5
227	65.2	31.9	95.9









#95
Naphthalene
Concen: 53.977 ug/l
RT: 15.13 min Scan# 4155
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tot Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.1	15.1
129	10.9	8.7	13.1

Manual Integrations
APPROVED

MMDadoda
3/22/2019 4:15:17 PM



#96
1,2,3-Trichlorobenzene
Concen: 54.499 ug/l
RT: 15.31 min Scan# 4211
Delta R.T. 0.00 min
Lab File: VN054505.D
Acq: 21 Mar 2019 11:57

Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.9	47.5	142.6
145	30.4	14.8	44.3

