

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071416.D
 Acq On : 21 Mar 2022 13:00
 Operator : JC\MD
 Sample : VN0321MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0321MBL01

Quant Time: Mar 21 16:24:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	615732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1034993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	935106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	286074	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	402294	47.672	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.340%
35) Dibromofluoromethane	8.016	113	309352	47.365	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.740%
50) Toluene-d8	10.439	98	1260084	47.918	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.840%
62) 4-Bromofluorobenzene	12.727	95	378026	42.453	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071416.D
Acq On : 21 Mar 2022 13:00
Operator : JC\MD
Sample : VN0321MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0321MBL01

Quant Time: Mar 21 16:24:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

