

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071429.D
 Acq On : 21 Mar 2022 19:41
 Operator : JC\MD
 Sample : N1985-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VACLOT-SOIL-2

Quant Time: Mar 22 06:22:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	764176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1296552	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1248885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	451785	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	497049	47.459	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.920%
35) Dibromofluoromethane	8.022	113	372643	45.546	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.100%
50) Toluene-d8	10.439	98	1612136	48.939	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.727	95	596821	53.503	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.000%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	135284	26.547	ug/l	97
18) Methyl Acetate	4.857	43	31707	2.765	ug/l	93
20) Methylene Chloride	5.093	84	21402	2.662	ug/l	91
25) 2-Butanone	7.334	43	67944	8.277	ug/l	93
43) Isopropyl Acetate	8.557	43	94781	3.513	ug/l #	83
95) Naphthalene	15.504	128	19351	5.907	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071429.D
Acq On : 21 Mar 2022 19:41
Operator : JC\MD
Sample : N1985-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VACLOT-SOIL-2

Quant Time: Mar 22 06:22:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

