

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071438.D
 Acq On : 22 Mar 2022 00:07
 Operator : JC\MD
 Sample : VN0321WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0321WBL02

Quant Time: Mar 22 06:27:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	777732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1310141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1239280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	447168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	502552	47.148	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.300%
35) Dibromofluoromethane	8.016	113	382689	46.289	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.580%
50) Toluene-d8	10.439	98	1622912	48.755	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.500%
62) 4-Bromofluorobenzene	12.727	95	588807	52.237	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071438.D
Acq On : 22 Mar 2022 00:07
Operator : JC\MD
Sample : VN0321WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0321WBL02

Quant Time: Mar 22 06:27:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

