

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071442.D
 Acq On : 22 Mar 2022 01:43
 Operator : JC\MD
 Sample : N2015-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-09

Quant Time: Mar 22 06:29:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	788767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1320854	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1261315	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	464455	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	502407	46.475	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.940%
35) Dibromofluoromethane	8.022	113	384142	46.088	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.180%
50) Toluene-d8	10.439	98	1661020	49.495	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.980%
62) 4-Bromofluorobenzene	12.727	95	613938	54.025	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.040%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.363	59	123847	58.318	ug/l	97
18) Methyl Acetate	4.863	43	9303	1.122	ug/l	100
19) Methyl tert-butyl Ether	5.616	73	127726	4.302	ug/l #	48
22) Diisopropyl ether	6.492	45	26789	0.791	ug/l #	79
95) Naphthalene	15.498	128	6981	5.362	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071442.D
Acq On : 22 Mar 2022 01:43
Operator : JC\MD
Sample : N2015-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-09

Quant Time: Mar 22 06:29:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

