

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071443.D
 Acq On : 22 Mar 2022 02:08
 Operator : JC\MD
 Sample : N2015-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2022-03-17

Quant Time: Mar 22 06:30:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	784378	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1327552	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1241204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	422049	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	523068	48.657	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.320%
35) Dibromofluoromethane	8.016	113	391367	46.717	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.440%
50) Toluene-d8	10.439	98	1642162	48.686	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.380%
62) 4-Bromofluorobenzene	12.727	95	584351	51.162	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.320%
Target Compounds						
18) Methyl Acetate	4.869	43	9909	1.169	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071443.D
Acq On : 22 Mar 2022 02:08
Operator : JC\MD
Sample : N2015-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2022-03-17

Quant Time: Mar 22 06:30:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

