

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071444.D
 Acq On : 22 Mar 2022 02:32
 Operator : JC\MD
 Sample : N2015-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-2022-03-17

Quant Time: Mar 22 06:30:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	749842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1278166	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1212318	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	433963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	483347	47.033	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.060%		
35) Dibromofluoromethane	8.016	113	370985	45.996	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.000%		
50) Toluene-d8	10.439	98	1555636	47.903	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.800%		
62) 4-Bromofluorobenzene	12.727	95	576588	52.432	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	104.860%		
Target Compounds						
16) Acetone	4.293	43	9509	1.902	ug/l	97
18) Methyl Acetate	4.857	43	9899	1.200	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071444.D
Acq On : 22 Mar 2022 02:32
Operator : JC\MD
Sample : N2015-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-2022-03-17

Quant Time: Mar 22 06:30:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

