

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071446.D
 Acq On : 22 Mar 2022 03:20
 Operator : JC\MD
 Sample : N2017-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Mar 22 06:31:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	722748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1227587	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1139606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	401725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	472288	47.679	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.360%
35) Dibromofluoromethane	8.016	113	358786	46.316	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.640%
50) Toluene-d8	10.439	98	1502968	48.188	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.380%
62) 4-Bromofluorobenzene	12.727	95	538777	51.013	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.020%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	17359	3.602	ug/l #	80
18) Methyl Acetate	4.857	43	11675	1.364	ug/l #	88
27) cis-1,2-Dichloroethene	7.316	96	7547	0.802	ug/l	84
30) Chloroform	7.822	83	10228	0.654	ug/l	94
44) Trichloroethene	9.210	130	5328	0.604	ug/l	72
52) Toluene	10.504	92	5797	0.252	ug/l	76
64) Tetrachloroethene	10.974	164	9965	1.254	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071446.D
Acq On : 22 Mar 2022 03:20
Operator : JC\MD
Sample : N2017-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Mar 22 06:31:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

