

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071451.D
 Acq On : 22 Mar 2022 05:21
 Operator : JC\MD
 Sample : N2017-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-3-17-22

Quant Time: Mar 22 06:33:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	746763	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1273129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1236945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	449787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	496214	48.484	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.960%
35) Dibromofluoromethane	8.016	113	370548	46.123	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.240%
50) Toluene-d8	10.439	98	1582858	48.934	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.727	95	589629	53.831	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.660%
Target Compounds						
16) Acetone	4.293	43	6925	1.391	ug/l #	82
18) Methyl Acetate	4.851	43	10119	1.220	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071451.D
Acq On : 22 Mar 2022 05:21
Operator : JC\MD
Sample : N2017-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-3-17-22

Quant Time: Mar 22 06:33:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

