

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071455.D
 Acq On : 22 Mar 2022 06:57
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 08:09:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.548	0.460	16.1	79	0.00
3 P	Chloromethane	0.822	0.657	20.1	80	0.00
4 C	Vinyl Chloride	0.672	0.618	8.0#	85	0.00
5 T	Bromomethane	0.335	0.302	9.9	79	-0.01
6 T	Chloroethane	0.407	0.419	-2.9	91	-0.01
7 T	Trichlorofluoromethane	0.884	0.835	5.5	88	0.00
8 T	Diethyl Ether	0.378	0.388	-2.6	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.512	0.483	5.7	88	0.00
10 T	Methyl Iodide	0.723	0.674	6.8	83	0.00
11 T	Tert butyl alcohol	0.135	0.152	-12.6	105	0.00
12 CM	1,1-Dichloroethene	0.501	0.478	4.6#	88	0.00
13 T	Acrolein	0.142	0.119	16.2	82	0.00
14 T	Allyl chloride	1.090	1.062	2.6	89	0.00
15 T	Acrylonitrile	0.400	0.425	-6.2	96	0.00
16 T	Acetone	0.333	0.362	-8.7	102	0.00
17 T	Carbon Disulfide	1.427	1.140	20.1	74	0.00
18 T	Methyl Acetate	0.979	1.005	-2.7	101	0.00
19 T	Methyl tert-butyl Ether	1.882	2.002	-6.4	96	0.00
20 T	Methylene Chloride	0.661	0.604	8.6	89	0.00
21 T	trans-1,2-Dichloroethene	0.548	0.519	5.3	87	0.00
22 T	Diisopropyl ether	2.148	2.258	-5.1	95	0.00
23 T	Vinyl Acetate	1.818	1.966	-8.1	94	0.00
24 P	1,1-Dichloroethane	1.135	1.145	-0.9	93	0.00
25 T	2-Butanone	0.537	0.594	-10.6	100	0.00
26 T	2,2-Dichloropropane	0.876	0.679	22.5	70	0.00
27 T	cis-1,2-Dichloroethene	0.651	0.654	-0.5	93	0.00
28 T	Bromochloromethane	0.505	0.439	13.1	79	0.00
29 T	Tetrahydrofuran	0.372	0.398	-7.0	97	0.00
30 C	Chloroform	1.083	1.113	-2.8#	95	0.00
31 T	Cyclohexane	1.094	0.993	9.2	86	0.00
32 T	1,1,1-Trichloroethane	0.922	0.943	-2.3	93	0.00
33 S	1,2-Dichloroethane-d4	0.685	0.617	9.9	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.316	0.276	12.7	87	0.00
36 T	1,1-Dichloropropene	0.506	0.468	7.5	89	0.00
37 T	Ethyl Acetate	0.696	0.671	3.6	92	0.00
38 T	Carbon Tetrachloride	0.489	0.462	5.5	90	0.00
39 T	Methylcyclohexane	0.595	0.539	9.4	84	0.00
40 TM	Benzene	1.523	1.454	4.5	91	0.00
41 T	Methacrylonitrile	0.340	0.336	1.2	94	0.00
42 TM	1,2-Dichloroethane	0.530	0.525	0.9	93	0.00
43 T	Isopropyl Acetate	1.040	1.049	-0.9	99	0.00
44 TM	Trichloroethene	0.359	0.343	4.5	90	0.00
45 C	1,2-Dichloropropane	0.409	0.405	1.0#	93	0.00
46 T	Dibromomethane	0.252	0.246	2.4	92	0.00
47 T	Bromodichloromethane	0.509	0.508	0.2	94	0.00
48 T	Methyl methacrylate	0.482	0.472	2.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071455.D
 Acq On : 22 Mar 2022 06:57
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 22 08:09:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	101	0.00
50 S	Toluene-d8	1.270	1.077	15.2	82	0.00
51 T	4-Methyl-2-Pentanone	0.646	0.700	-8.4	100	0.00
52 CM	Toluene	0.938	0.910	3.0#	89	0.00
53 T	t-1,3-Dichloropropene	0.561	0.544	3.0	89	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.593	3.3	89	0.00
55 T	1,1,2-Trichloroethane	0.376	0.369	1.9	95	0.00
56 T	Ethyl methacrylate	0.614	0.660	-7.5	98	0.00
57 T	1,3-Dichloropropane	0.675	0.669	0.9	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.101	0.141	-39.6#	119	0.00
59 T	2-Hexanone	0.441	0.514	-16.6	104	0.00
60 T	Dibromochloromethane	0.373	0.385	-3.2	93	0.00
61 T	1,2-Dibromoethane	0.384	0.374	2.6	92	0.00
62 S	4-Bromofluorobenzene	0.430	0.379	11.9	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.349	0.303	13.2	84	0.00
65 PM	Chlorobenzene	1.074	1.031	4.0	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.389	1.3	93	0.00
67 C	Ethyl Benzene	1.899	1.885	0.7#	91	0.00
68 T	m/p-Xylenes	0.695	0.700	-0.7	91	0.00
69 T	o-Xylene	0.696	0.710	-2.0	92	0.00
70 T	Styrene	1.078	1.155	-7.1	95	0.00
71 P	Bromoform	0.299	0.315	-5.4	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	4.820	4.515	6.3	93	0.00
74 T	N-amyl acetate	1.812	1.960	-8.2	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.655	1.562	5.6	98	0.00
76 T	1,2,3-Trichloropropane	1.407	1.209	14.1	93	0.00
77 T	Bromobenzene	1.124	1.043	7.2	94	0.00
78 T	n-propylbenzene	5.172	4.998	3.4	93	0.00
79 T	2-Chlorotoluene	3.320	3.077	7.3	93	0.00
80 T	1,3,5-Trimethylbenzene	3.750	3.553	5.3	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.463	0.434	6.3	89	0.00
82 T	4-Chlorotoluene	2.957	2.890	2.3	94	0.00
83 T	tert-Butylbenzene	3.347	3.306	1.2	95	0.00
84 T	1,2,4-Trimethylbenzene	3.592	3.513	2.2	93	0.00
85 T	sec-Butylbenzene	4.448	4.384	1.4	92	0.00
86 T	p-Isopropyltoluene	3.498	3.515	-0.5	92	0.00
87 T	1,3-Dichlorobenzene	1.826	1.709	6.4	93	0.00
88 T	1,4-Dichlorobenzene	1.785	1.645	7.8	94	0.00
89 T	n-Butylbenzene	2.736	2.626	4.0	91	0.00
90 T	Hexachloroethane	0.696	0.677	2.7	94	0.00
91 T	1,2-Dichlorobenzene	1.756	1.674	4.7	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.280	0.296	-5.7	106	0.00
93 T	1,2,4-Trichlorobenzene	0.797	0.807	-1.3	99	0.00
94 T	Hexachlorobutadiene	0.493	0.437	11.4	86	0.00
95 T	Naphthalene	2.115	2.626	-24.2	123	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071455.D
Acq On : 22 Mar 2022 06:57
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 22 08:09:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.775	0.817	-5.4	103	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6