

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
 Data File : VN071438.D
 Acq On : 22 Mar 2022 00:07
 Operator : JC\MD
 Sample : VN0321WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0321WBL02

Quant Time: Mar 22 06:27:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	777732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1310141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1239280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	447168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	502552	47.148	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.300%
35) Dibromofluoromethane	8.016	113	382689	46.289	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.580%
50) Toluene-d8	10.439	98	1622912	48.755	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.500%
62) 4-Bromofluorobenzene	12.727	95	588807	52.237	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.480%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032122\
Data File : VN071438.D
Acq On : 22 Mar 2022 00:07
Operator : JC\MD
Sample : VN0321WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0321WBL02

Quant Time: Mar 22 06:27:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.081 min Scan# 1036
Delta R.T. -0.005 min
Lab File: VN071438.D
Acq: 22 Mar 2022 00:07




Instrument :
MSVOA_N
ClientSampleId :
VN0321WBL02


Tgt Ion:168 Resp: 777732
Ion Ratio Lower Upper
168 100
99 56.4 41.3 61.9



#33
1,2-Dichloroethane-d4
Concen: 47.148 ug/l
RT: 8.434 min Scan# 1096
Delta R.T. -0.006 min
Lab File: VN071438.D
Acq: 22 Mar 2022 00:07

Tgt Ion: 65 Resp: 502552
Ion Ratio Lower Upper
65 100
67 52.9 0.0 103.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.963 min Scan# 1186
Delta R.T. -0.000 min
Lab File: VN071438.D
Acq: 22 Mar 2022 00:07




Instrument :
MSVOA_N
ClientSampleId :
VN0321WBL02


Tgt Ion:114 Resp: 1310141
Ion Ratio Lower Upper
114 100
63 22.3 0.0 37.8
88 16.1 0.0 27.4



#35
Dibromofluoromethane
Concen: 46.289 ug/l
RT: 8.016 min Scan# 1025
Delta R.T. -0.006 min
Lab File: VN071438.D
Acq: 22 Mar 2022 00:07

Tgt Ion:113 Resp: 382689
Ion Ratio Lower Upper
113 100
111 102.1 82.2 123.2
192 20.0 16.9 25.3









#50
Toluene-d8
Concen: 48.755 ug/l
RT: 10.439 min Scan# 1437
Delta R.T. 0.000 min
Lab File: VN071438.D
Acq: 22 Mar 2022 00:07




Instrument :
MSVOA_N
ClientSampleId :
VN0321WBL02


Tgt Ion: 98 Resp: 1622912
Ion Ratio Lower Upper
98 100
100 65.4 50.4 75.6



#62
4-Bromofluorobenzene
Concen: 52.237 ug/l
RT: 12.727 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN071438.D
Acq: 22 Mar 2022 00:07

Tgt Ion: 95 Resp: 588807
Ion Ratio Lower Upper
95 100
174 78.8 0.0 187.2
176 76.1 0.0 181.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.739 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN071438.D
Acq: 22 Mar 2022 00:07

Instrument :
MSVOA_N
ClientSampleId :
VN0321WBL02

Tgt Ion:117 Resp: 1239280
Ion Ratio Lower Upper
117 100
82 57.2 42.4 63.6
119 31.4 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.669 min Scan# 1986
Delta R.T. 0.001 min
Lab File: VN071438.D
Acq: 22 Mar 2022 00:07

Tgt Ion:152 Resp: 447168
Ion Ratio Lower Upper
152 100
115 59.7 27.5 82.5
150 158.6 0.0 344.4

