

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
 Data File : VN077083.D
 Acq On : 21 Mar 2023 11:29
 Operator : JC\MD
 Sample : VN0321MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0321MBL01

Quant Time: Mar 21 23:13:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	375472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	673951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	584089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	214175	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	279071	46.531	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.060%
35) Dibromofluoromethane	8.172	113	214676	45.484	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.960%
50) Toluene-d8	10.572	98	812040	48.165	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.320%
62) 4-Bromofluorobenzene	12.854	95	260620	43.746	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
Data File : VN077083.D
Acq On : 21 Mar 2023 11:29
Operator : JC\MD
Sample : VN0321MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0321MBL01

Quant Time: Mar 21 23:13:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

