

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
 Data File : VN077087.D
 Acq On : 21 Mar 2023 13:52
 Operator : JC\MD
 Sample : 01941-04DL 4X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT180D1-20230313DL

Quant Time: Mar 21 23:16:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	400950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	728016	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	636741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	234653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	296629	46.316	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.640%
35) Dibromofluoromethane	8.172	113	228931	44.903	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.800%
50) Toluene-d8	10.572	98	882503	48.457	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.920%
62) 4-Bromofluorobenzene	12.854	95	286938	44.587	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.180%
Target Compounds						
					Qvalue	
30) Chloroform	7.972	83	2524	0.233	ug/l	98
44) Trichloroethene	9.354	130	230808	41.607	ug/l	94
95) Naphthalene	15.642	128	8000	2.405	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
Data File : VN077087.D
Acq On : 21 Mar 2023 13:52
Operator : JC\MD
Sample : 01941-04DL 4X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT180D1-20230313DL

Quant Time: Mar 21 23:16:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

