

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
 Data File : VN077090.D
 Acq On : 21 Mar 2023 15:05
 Operator : JC\MD
 Sample : 02001-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IRV-1

Quant Time: Mar 21 23:17:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	435291	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	805430	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	714501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	262479	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	328684	47.272	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.540%
35) Dibromofluoromethane	8.172	113	252102	44.695	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.380%
50) Toluene-d8	10.572	98	989524	49.111	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.220%
62) 4-Bromofluorobenzene	12.854	95	319359	44.855	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.700%
Target Compounds						
					Qvalue	
16) Acetone	4.449	43	44906	14.495	ug/l	90
20) Methylene Chloride	5.290	84	20694	3.078	ug/l	94
37) Ethyl Acetate	7.572	43	84729	8.989	ug/l	99
43) Isopropyl Acetate	8.695	43	461461	31.338	ug/l #	84
95) Naphthalene	15.654	128	1639	2.002	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
Data File : VN077090.D
Acq On : 21 Mar 2023 15:05
Operator : JC\MD
Sample : 02001-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IRV-1

Quant Time: Mar 21 23:17:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

