

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
 Data File : VN077091.D
 Acq On : 21 Mar 2023 15:29
 Operator : JC\MD
 Sample : 01975-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CHAMBER-B-72-WEST

Quant Time: Mar 21 23:18:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	386497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	717599	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	660054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	251959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	302927	49.068	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.140%		
35) Dibromofluoromethane	8.172	113	224854	44.743	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.480%		
50) Toluene-d8	10.572	98	903180	50.312	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.620%		
62) 4-Bromofluorobenzene	12.854	95	294159	46.372	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.740%		
Target Compounds						
						Qvalue
16) Acetone	4.443	43	129321	47.012	ug/l	97
20) Methylene Chloride	5.290	84	15396	2.579	ug/l	93
25) 2-Butanone	7.495	43	45618	11.336	ug/l	97
37) Ethyl Acetate	7.566	43	77536	9.232	ug/l	99
43) Isopropyl Acetate	8.695	43	366704	27.951	ug/l #	89
52) Toluene	10.636	92	2443037	171.883	ug/l	99
65) Chlorobenzene	11.895	112	24301	1.600	ug/l	93
95) Naphthalene	15.642	128	6495	2.285	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
Data File : VN077091.D
Acq On : 21 Mar 2023 15:29
Operator : JC\MD
Sample : 01975-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CHAMBER-B-72-WEST

Quant Time: Mar 21 23:18:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

