

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
 Data File : VN077094.D
 Acq On : 21 Mar 2023 16:47
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 21 23:19:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.566	0.357	36.9#	73	0.00
3 P	Chloromethane	0.504	0.321	36.3#	67	0.00
4 C	Vinyl Chloride	0.544	0.407	25.2#	77	0.00
5 T	Bromomethane	0.367	0.320	12.8	85	0.00
6 T	Chloroethane	0.344	0.351	-2.0	110	0.00
7 T	Trichlorofluoromethane	1.170	1.036	11.5	91	0.00
8 T	Diethyl Ether	0.440	0.418	5.0	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.664	0.566	14.8	90	0.00
10 T	Methyl Iodide	0.747	0.671	10.2	87	0.00
11 T	Tert butyl alcohol	0.116	0.118	-1.7	99	0.00
12 CM	1,1-Dichloroethene	0.607	0.537	11.5#	89	0.00
13 T	Acrolein	0.144	0.139	3.5	97	0.00
14 T	Allyl chloride	1.111	1.013	8.8	89	0.00
15 T	Acrylonitrile	0.391	0.376	3.8	95	0.00
16 T	Acetone	0.356	0.332	6.7	97	0.00
17 T	Carbon Disulfide	1.770	1.463	17.3	85	0.00
18 T	Methyl Acetate	1.342	1.304	2.8	98	0.00
19 T	Methyl tert-butyl Ether	2.179	2.099	3.7	95	0.00
20 T	Methylene Chloride	0.772	0.691	10.5	96	0.00
21 T	trans-1,2-Dichloroethene	0.658	0.599	9.0	92	0.00
22 T	Diisopropyl ether	2.435	2.413	0.9	95	0.00
23 T	Vinyl Acetate	1.536	1.534	0.1	95	0.00
24 P	1,1-Dichloroethane	1.393	1.280	8.1	94	0.00
25 T	2-Butanone	0.521	0.513	1.5	95	0.00
26 T	2,2-Dichloropropane	1.048	0.966	7.8	90	0.00
27 T	cis-1,2-Dichloroethene	0.754	0.723	4.1	94	0.00
28 T	Bromochloromethane	0.605	0.594	1.8	104	0.00
29 T	Tetrahydrofuran	0.337	0.342	-1.5	96	0.00
30 C	Chloroform	1.352	1.257	7.0#	95	0.00
31 T	Cyclohexane	1.307	1.095	16.2	87	0.00
32 T	1,1,1-Trichloroethane	1.156	1.053	8.9	91	0.00
33 S	1,2-Dichloroethane-d4	0.799	0.830	-3.9	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.350	0.335	4.3	100	0.00
36 T	1,1-Dichloropropene	0.547	0.499	8.8	89	0.00
37 T	Ethyl Acetate	0.585	0.566	3.2	97	0.00
38 T	Carbon Tetrachloride	0.533	0.488	8.4	91	0.00
39 T	Methylcyclohexane	0.640	0.593	7.3	86	0.00
40 TM	Benzene	1.659	1.508	9.1	92	0.00
41 T	Methacrylonitrile	0.315	0.323	-2.5	103	0.00
42 TM	1,2-Dichloroethane	0.599	0.563	6.0	96	0.00
43 T	Isopropyl Acetate	0.914	0.940	-2.8	98	0.00
44 TM	Trichloroethene	0.381	0.325	14.7	88	0.00
45 C	1,2-Dichloropropane	0.441	0.413	6.3#	94	0.00
46 T	Dibromomethane	0.278	0.258	7.2	95	0.00
47 T	Bromodichloromethane	0.583	0.544	6.7	95	0.00
48 T	Methyl methacrylate	0.437	0.454	-3.9	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
 Data File : VN077094.D
 Acq On : 21 Mar 2023 16:47
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 21 23:19:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 20 08:50:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	99	0.00
50 S	Toluene-d8	1.251	1.261	-0.8	98	0.00
51 T	4-Methyl-2-Pentanone	0.569	0.582	-2.3	97	0.00
52 CM	Toluene	0.990	0.924	6.7#	90	0.00
53 T	t-1,3-Dichloropropene	0.571	0.580	-1.6	93	0.00
54 T	cis-1,3-Dichloropropene	0.644	0.641	0.5	95	0.00
55 T	1,1,2-Trichloroethane	0.391	0.375	4.1	96	0.00
56 T	Ethyl methacrylate	0.587	0.606	-3.2	93	0.00
57 T	1,3-Dichloropropane	0.715	0.673	5.9	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.237	5.2	87	0.00
59 T	2-Hexanone	0.413	0.423	-2.4	96	0.00
60 T	Dibromochloromethane	0.392	0.379	3.3	93	0.00
61 T	1,2-Dibromoethane	0.387	0.367	5.2	95	0.00
62 S	4-Bromofluorobenzene	0.442	0.465	-5.2	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.336	0.287	14.6	88	0.00
65 PM	Chlorobenzene	1.150	1.044	9.2	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.426	0.385	9.6	93	0.00
67 C	Ethyl Benzene	2.090	2.000	4.3#	91	0.00
68 T	m/p-Xylenes	0.775	0.735	5.2	90	0.00
69 T	o-Xylene	0.770	0.735	4.5	92	0.00
70 T	Styrene	1.199	1.219	-1.7	92	0.00
71 P	Bromoform	0.288	0.280	2.8	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	4.481	4.166	7.0	91	0.00
74 T	N-amyl acetate	1.939	1.951	-0.6	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.623	1.384	14.7	95	0.00
76 T	1,2,3-Trichloropropane	1.293	1.286	0.5	98	0.00
77 T	Bromobenzene	1.004	0.887	11.7	94	0.00
78 T	n-propylbenzene	5.173	4.982	3.7	91	0.00
79 T	2-Chlorotoluene	3.296	3.002	8.9	91	0.00
80 T	1,3,5-Trimethylbenzene	3.741	3.560	4.8	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.470	0.494	-5.1	105	0.00
82 T	4-Chlorotoluene	3.134	2.959	5.6	94	0.00
83 T	tert-Butylbenzene	3.141	2.921	7.0	90	0.00
84 T	1,2,4-Trimethylbenzene	3.733	3.544	5.1	91	0.00
85 T	sec-Butylbenzene	4.468	4.369	2.2	91	0.00
86 T	p-Isopropyltoluene	3.588	3.552	1.0	92	0.00
87 T	1,3-Dichlorobenzene	1.897	1.672	11.9	93	0.00
88 T	1,4-Dichlorobenzene	1.887	1.642	13.0	92	0.00
89 T	n-Butylbenzene	3.148	3.169	-0.7	91	0.00
90 T	Hexachloroethane	0.730	0.640	12.3	92	0.00
91 T	1,2-Dichlorobenzene	1.891	1.661	12.2	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.301	0.280	7.0	92	0.00
93 T	1,2,4-Trichlorobenzene	0.855	0.835	2.3	87	0.00
94 T	Hexachlorobutadiene	0.471	0.407	13.6	87	0.00
95 T	Naphthalene	2.729	2.920	-7.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032123\
Data File : VN077094.D
Acq On : 21 Mar 2023 16:47
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 21 23:19:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 20 08:50:30 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.858	0.831	3.1	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6