

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086040.D
 Acq On : 21 Mar 2025 15:17
 Operator : JC\MD
 Sample : PB167234TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167234TB

Quant Time: Mar 21 22:58:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	133448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	243178	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	238825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	82133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	105034	57.479	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 114.960%		
35) Dibromofluoromethane	8.165	113	88268	52.003	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.000%		
50) Toluene-d8	10.565	98	322533	52.357	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 104.720%		
62) 4-Bromofluorobenzene	12.847	95	95610	43.520	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 87.040%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	26092	46.737	ug/l	97
20) Methylene Chloride	5.277	84	10771	6.595	ug/l	91
43) Isopropyl Acetate	8.688	43	88364	21.627	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086040.D
Acq On : 21 Mar 2025 15:17
Operator : JC\MD
Sample : PB167234TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167234TB

Quant Time: Mar 21 22:58:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

