

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086041.D
 Acq On : 21 Mar 2025 15:41
 Operator : JC\MD
 Sample : PB167234ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167234ZHE#01

Quant Time: Mar 21 22:59:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	145094	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	263894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	98136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112226	56.485	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 112.960%		
35) Dibromofluoromethane	8.165	113	95750	51.983	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.960%		
50) Toluene-d8	10.565	98	353841	52.931	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 105.860%		
62) 4-Bromofluorobenzene	12.847	95	105907	44.423	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 88.840%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	26476	43.618	ug/l	99
20) Methylene Chloride	5.277	84	9759	5.496	ug/l #	82
25) 2-Butanone	7.494	43	6485	7.404	ug/l #	76
43) Isopropyl Acetate	8.688	43	92187	20.791	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086041.D
Acq On : 21 Mar 2025 15:41
Operator : JC\MD
Sample : PB167234ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167234ZHE#01

Quant Time: Mar 21 22:59:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

