

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086043.D
 Acq On : 21 Mar 2025 16:29
 Operator : JC\MD
 Sample : PB167234ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167234ZHE#03

Quant Time: Mar 21 23:00:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	149330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	272825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	257981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113434	55.473	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.940%			
35) Dibromofluoromethane	8.165	113	96399	50.622	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.240%			
50) Toluene-d8	10.565	98	351782	50.900	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.800%			
62) 4-Bromofluorobenzene	12.847	95	101069	41.006	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 82.020%			
Target Compounds						
16) Acetone	4.430	43	24547	39.293	ug/l	99
20) Methylene Chloride	5.271	84	10396	5.688	ug/l #	96
25) 2-Butanone	7.489	43	5873	6.515	ug/l	98
43) Isopropyl Acetate	8.688	43	95815	20.902	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086043.D
Acq On : 21 Mar 2025 16:29
Operator : JC\MD
Sample : PB167234ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167234ZHE#03

Quant Time: Mar 21 23:00:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

