

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086045.D
 Acq On : 21 Mar 2025 17:18
 Operator : JC\MD
 Sample : PB167234ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167234ZHE#05

Quant Time: Mar 21 23:01:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	139623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	256399	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255466	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	94489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	108915	56.966	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.940%
35) Dibromofluoromethane	8.165	113	94617	52.869	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.740%
50) Toluene-d8	10.565	98	341159	52.525	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.060%
62) 4-Bromofluorobenzene	12.847	95	101386	43.770	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.540%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	22223	38.046	ug/l	95
20) Methylene Chloride	5.271	84	10069	5.893	ug/l #	91
25) 2-Butanone	7.483	43	5462	6.481	ug/l	88
43) Isopropyl Acetate	8.688	43	100309	23.284	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086045.D
Acq On : 21 Mar 2025 17:18
Operator : JC\MD
Sample : PB167234ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167234ZHE#05

Quant Time: Mar 21 23:01:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

