

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086049.D
 Acq On : 21 Mar 2025 18:54
 Operator : JC\MD
 Sample : PB167234ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167234ZHE#09

Quant Time: Mar 21 23:03:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	129886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	251636	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90956	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	105914	59.550	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 119.100%			
35) Dibromofluoromethane	8.165	113	89439	50.922	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.840%			
50) Toluene-d8	10.565	98	325371	51.043	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.080%			
62) 4-Bromofluorobenzene	12.847	95	96802	42.582	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 85.160%			
Target Compounds						
16) Acetone	4.424	43	22542	41.485	ug/l	88
20) Methylene Chloride	5.283	84	10166	6.395	ug/l	91
43) Isopropyl Acetate	8.688	43	108566	25.678	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086049.D
Acq On : 21 Mar 2025 18:54
Operator : JC\MD
Sample : PB167234ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167234ZHE#09

Quant Time: Mar 21 23:03:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

