

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086050.D
 Acq On : 21 Mar 2025 19:18
 Operator : JC\MD
 Sample : PB167234ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167234ZHE#10

Quant Time: Mar 21 23:03:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	135704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	260868	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	257655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	93569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109916	59.150	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	118.300%	
35) Dibromofluoromethane	8.165	113	92595	50.853	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.700%	
50) Toluene-d8	10.565	98	340992	51.600	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.200%	
62) 4-Bromofluorobenzene	12.847	95	101706	43.156	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	86.320%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	22900	40.337	ug/l	96
20) Methylene Chloride	5.277	84	10189	6.135	ug/l	95
43) Isopropyl Acetate	8.683	43	111737	25.493	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086050.D
Acq On : 21 Mar 2025 19:18
Operator : JC\MD
Sample : PB167234ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167234ZHE#10

Quant Time: Mar 21 23:03:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

