

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086052.D
 Acq On : 21 Mar 2025 20:06
 Operator : JC\MD
 Sample : PB167234ZHE#12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167234ZHE#12

Quant Time: Mar 21 23:04:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	146866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	285919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	271974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	97362	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	115936	57.648	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.300%			
35) Dibromofluoromethane	8.165	113	100760	50.489	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.980%			
50) Toluene-d8	10.565	98	364624	50.342	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.680%			
62) 4-Bromofluorobenzene	12.847	95	107070	41.451	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 82.900%			
Target Compounds					Qvalue	
16) Acetone	4.424	43	25110	40.869	ug/l	92
20) Methylene Chloride	5.277	84	10092	5.615	ug/l #	92
43) Isopropyl Acetate	8.688	43	123813	25.773	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086052.D
Acq On : 21 Mar 2025 20:06
Operator : JC\MD
Sample : PB167234ZHE#12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167234ZHE#12

Quant Time: Mar 21 23:04:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

