

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086053.D
 Acq On : 21 Mar 2025 20:30
 Operator : JC\MD
 Sample : PB167234ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167234ZHE#13

Quant Time: Mar 21 23:04:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	142056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	254743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	90983	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110134	56.617	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.240%			
35) Dibromofluoromethane	8.165	113	93717	49.367	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.740%			
50) Toluene-d8	10.565	98	344197	49.957	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.920%			
62) 4-Bromofluorobenzene	12.847	95	99813	40.622	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 81.240%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	24396	41.051	ug/l	100
20) Methylene Chloride	5.277	84	9626	5.537	ug/l	97
43) Isopropyl Acetate	8.688	43	120540	26.378	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086053.D
Acq On : 21 Mar 2025 20:30
Operator : JC\MD
Sample : PB167234ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167234ZHE#13

Quant Time: Mar 21 23:04:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

