

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086059.D
 Acq On : 21 Mar 2025 22:55
 Operator : JC\MD
 Sample : Q1606-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT3025

Quant Time: Mar 21 23:59:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	134777	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	260087	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	258020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	97446	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110383	59.810	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.620%
35) Dibromofluoromethane	8.159	113	91986	50.671	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.340%
50) Toluene-d8	10.565	98	336768	51.114	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.220%
62) 4-Bromofluorobenzene	12.847	95	101074	43.016	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.040%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	24763	43.919	ug/l	96
20) Methylene Chloride	5.271	84	9512	5.767	ug/l	91
40) Benzene	8.606	78	47260	6.298	ug/l	99
43) Isopropyl Acetate	8.688	43	67458	15.437	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086059.D
Acq On : 21 Mar 2025 22:55
Operator : JC\MD
Sample : Q1606-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT3025

Quant Time: Mar 21 23:59:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

