

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086061.D
 Acq On : 21 Mar 2025 23:43
 Operator : JC\MD
 Sample : Q1609-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-SCRN-01-G

Quant Time: Mar 22 00:00:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	140609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	254973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	98590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110762	57.526	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.060%			
35) Dibromofluoromethane	8.165	113	93335	50.927	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.860%			
50) Toluene-d8	10.565	98	337987	50.814	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.620%			
62) 4-Bromofluorobenzene	12.847	95	101958	42.982	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 85.960%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	27174	46.196	ug/l	98
20) Methylene Chloride	5.271	84	9766	5.675	ug/l #	95
27) cis-1,2-Dichloroethene	7.489	96	33966	18.986	ug/l	91
43) Isopropyl Acetate	8.688	43	35240	7.988	ug/l #	68
44) Trichloroethene	9.347	130	8180	4.166	ug/l	99
95) Naphthalene	15.635	128	5135	1.048	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086061.D
Acq On : 21 Mar 2025 23:43
Operator : JC\MD
Sample : Q1609-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-SCRN-01-G

Quant Time: Mar 22 00:00:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

