

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
 Data File : VN086062.D
 Acq On : 22 Mar 2025 00:07
 Operator : JC\MD
 Sample : Q1619-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Mar 22 01:06:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	136258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262021	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	86621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110936	59.456	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 118.920%	
35) Dibromofluoromethane	8.165	113	92605	50.635	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.260%	
50) Toluene-d8	10.565	98	332573	50.105	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.200%	
62) 4-Bromofluorobenzene	12.847	95	95814	40.477	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 80.960%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	29097	51.045	ug/l	94
20) Methylene Chloride	5.283	84	8925	5.352	ug/l	85
25) 2-Butanone	7.494	43	4203	5.110	ug/l	92
43) Isopropyl Acetate	8.688	43	94241	21.407	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032125\
Data File : VN086062.D
Acq On : 22 Mar 2025 00:07
Operator : JC\MD
Sample : Q1619-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Mar 22 01:06:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

