

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032219\
 Data File : VN054549.D
 Acq On : 22 Mar 2019 11:10
 Operator : JC/SP
 Sample : VN0322WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0322WBL01

Quant Time: Mar 22 12:29:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 14:42:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	643784	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	981757	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	799676	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	269061	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	393166	49.10	ug/l	0.00
Spiked Amount			Recovery	=	98.20%	
35) Dibromofluoromethane	7.59	113	303894	48.05	ug/l	0.00
Spiked Amount			Recovery	=	96.10%	
50) Toluene-d8	10.09	98	1165275	49.01	ug/l	0.00
Spiked Amount			Recovery	=	98.02%	
62) 4-Bromofluorobenzene	12.40	95	345025	42.64	ug/l	0.00
Spiked Amount			Recovery	=	85.28%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054549.D
Acq On : 22 Mar 2019 11:10
Operator : JC/SP
Sample : VN0322WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0322WBL01

Quant Time: Mar 22 12:29:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 14:42:09 2019
Response via : Initial Calibration

