

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032219\
 Data File : VN054553.D
 Acq On : 22 Mar 2019 13:03
 Operator : JC/SP
 Sample : K2004-09
 Misc : 5.24g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-22(7-8)

Quant Time: Mar 25 15:22:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	612144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	937469	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	755312	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	321688	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	364987	47.94	ug/l	0.00
Spiked Amount						
			Recovery	=	95.88%	
35) Dibromofluoromethane	7.59	113	274028	45.37	ug/l	0.00
Spiked Amount						
			Recovery	=	90.74%	
50) Toluene-d8	10.09	98	1128341	49.70	ug/l	0.00
Spiked Amount						
			Recovery	=	99.40%	
62) 4-Bromofluorobenzene	12.40	95	358700	46.43	ug/l	0.00
Spiked Amount						
			Recovery	=	92.86%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054553.D
Acq On : 22 Mar 2019 13:03
Operator : JC/SP
Sample : K2004-09
Misc : 5.24g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-22(7-8)

Quant Time: Mar 25 15:22:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

