

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032219\
 Data File : VN054555.D
 Acq On : 22 Mar 2019 13:58
 Operator : JC/SP
 Sample : K2019-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OUTFALL001-190319

Quant Time: Mar 23 03:37:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	665928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1025968	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	821495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	283625	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	456621	55.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.26%	
35) Dibromofluoromethane	7.59	113	350526	53.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.06%	
50) Toluene-d8	10.09	98	1352337	54.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.84%	
62) 4-Bromofluorobenzene	12.40	95	399950	47.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.60%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.82	96	29534	3.775	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054555.D
Acq On : 22 Mar 2019 13:58
Operator : JC/SP
Sample : K2019-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OUTFALL001-190319

Quant Time: Mar 23 03:37:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

