

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032219\
 Data File : VN054559.D
 Acq On : 22 Mar 2019 15:47
 Operator : JC/SP
 Sample : K2043-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-QC-190319-TB2

Quant Time: Mar 23 03:47:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	617788	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	984772	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	789216	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	262383	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	435694	56.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.40%	
35) Dibromofluoromethane	7.59	113	336769	53.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.16%	
50) Toluene-d8	10.09	98	1256812	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
62) 4-Bromofluorobenzene	12.40	95	367012	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054559.D
Acq On : 22 Mar 2019 15:47
Operator : JC/SP
Sample : K2043-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-QC-190319-TB2

Quant Time: Mar 23 03:47:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

