

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032219\
 Data File : VN054564.D
 Acq On : 22 Mar 2019 18:05
 Operator : JC/SP
 Sample : K1965-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-HN-MW24IR-20190315

Quant Time: Mar 23 03:59:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	620532	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	991302	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	802083	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	272809	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	432942	56.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.18%	
35) Dibromofluoromethane	7.59	113	336998	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
50) Toluene-d8	10.09	98	1262898	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
62) 4-Bromofluorobenzene	12.40	95	373184	45.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.36%	
Target Compounds						
7) Trichlorofluoromethane	3.03	101	8517	0.865	ug/l	99
19) Methyl tert-butyl Ether	5.05	73	13891	0.795	ug/l	95
44) Trichloroethene	8.84	130	20044	2.720	ug/l	88
64) Tetrachloroethene	10.63	164	16056	2.413	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054564.D
Acq On : 22 Mar 2019 18:05
Operator : JC/SP
Sample : K1965-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-HN-MW24IR-20190315

Quant Time: Mar 23 03:59:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

