

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032219\
 Data File : VN054566.D
 Acq On : 22 Mar 2019 19:00
 Operator : JC/SP
 Sample : K1965-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPS1-TT-MW303I2-20190315

Quant Time: Mar 23 04:06:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	612815	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	955161	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	770242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	265917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	422860	55.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.96%	
35) Dibromofluoromethane	7.59	113	329446	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
50) Toluene-d8	10.09	98	1227569	53.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.12%	
62) 4-Bromofluorobenzene	12.40	95	363742	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
Target Compounds						
44) Trichloroethene	8.84	130	11362	1.600	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054566.D
Acq On : 22 Mar 2019 19:00
Operator : JC/SP
Sample : K1965-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPS1-TT-MW303I2-20190315

Quant Time: Mar 23 04:06:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

