

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032219\
 Data File : VN054581.D
 Acq On : 23 Mar 2019 2:48
 Operator : JC/SP
 Sample : K2045-07
 Misc : 7.11g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022SB438-190318-(19-21)

Quant Time: Mar 23 05:22:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	545441	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	818173	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	710034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	290580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	344093	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
35) Dibromofluoromethane	7.59	113	250742	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
50) Toluene-d8	10.09	98	1007185	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
62) 4-Bromofluorobenzene	12.40	95	325092	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
Target Compounds						
4) Vinyl Chloride	2.19	62	10433	1.465	ug/l	98
27) cis-1,2-Dichloroethene	6.82	96	127806	19.946	ug/l	96
44) Trichloroethene	8.83	130	11435	1.880	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054581.D
Acq On : 23 Mar 2019 2:48
Operator : JC/SP
Sample : K2045-07
Misc : 7.11µ/5mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022SB438-190318-(19-21)

Quant Time: Mar 23 05:22:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

