

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032219\
 Data File : VN054584.D
 Acq On : 23 Mar 2019 4:10
 Operator : JC/SP
 Sample : K2045-10
 Misc : 5.38g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022SB435-190318-(20-22)

Quant Time: Mar 23 05:28:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	555641	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	873111	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	745018	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	317770	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	351314	50.83	ug/l	0.00
Spiked Amount			Recovery	=	101.66%	
35) Dibromofluoromethane	7.59	113	258701	45.99	ug/l	0.00
Spiked Amount			Recovery	=	91.98%	
50) Toluene-d8	10.09	98	1054926	49.89	ug/l	0.00
Spiked Amount			Recovery	=	99.78%	
62) 4-Bromofluorobenzene	12.40	95	340944	47.38	ug/l	0.00
Spiked Amount			Recovery	=	94.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.19	62	112720	15.535	ug/l	97
12) 1,1-Dichloroethene	3.71	96	35342	6.594	ug/l #	75
21) trans-1,2-Dichloroethene	5.02	96	31594	5.773	ug/l #	82
27) cis-1,2-Dichloroethene	6.82	96	1939725	297.164	ug/l	97
44) Trichloroethene	8.83	130	5514	0.850	ug/l	93
64) Tetrachloroethene	10.63	164	12148	1.966	ug/l	93
65) Chlorobenzene	11.43	112	10840	0.715	ug/l	93
87) 1,3-Dichlorobenzene	13.28	146	49644	4.717	ug/l	96
88) 1,4-Dichlorobenzene	13.36	146	171998	16.797	ug/l	95
91) 1,2-Dichlorobenzene	13.65	146	817938	80.240	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN032219\
Data File : VN054584.D
Acq On : 23 Mar 2019 4:10
Operator : JC/SP
Sample : K2045-10
Misc : 5.38g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022SB435-190318-(20-22)

Quant Time: Mar 23 05:28:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

