

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN032219\
 Data File : VN054591.D
 Acq On : 23 Mar 2019 7:23
 Operator : JC/SP
 Sample : K2045-17ME
 Misc : 5.13g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022SB444-190320-(14-16)ME

Quant Time: Mar 25 02:19:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	577876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	896109	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	768214	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	326760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	356445	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
35) Dibromofluoromethane	7.59	113	262147	45.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.82%	
50) Toluene-d8	10.09	98	1080570	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
62) 4-Bromofluorobenzene	12.40	95	353158	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
Target Compounds						
20) Methylene Chloride	4.53	84	17525	2.751	ug/l #	84
65) Chlorobenzene	11.43	112	127580	8.165	ug/l	97
87) 1,3-Dichlorobenzene	13.28	146	8824	0.815	ug/l #	70
88) 1,4-Dichlorobenzene	13.36	146	26021	2.471	ug/l #	78
91) 1,2-Dichlorobenzene	13.65	146	18917	1.805	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054591.D
Acq On : 23 Mar 2019 7:23
Operator : JC/SP
Sample : K2045-17ME
Misc : 5.13g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022SB444-190320-(14-16)ME

Quant Time: Mar 25 02:19:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

