

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN032219\
 Data File : VN054593.D
 Acq On : 23 Mar 2019 8:18
 Operator : JC/SP
 Sample : K2045-19
 Misc : 5.93g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP022SB444-190320-(23-25)

Quant Time: Mar 25 02:25:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 22 02:09:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	586895	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	911016	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	785042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	318544	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	360980	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
35) Dibromofluoromethane	7.59	113	272749	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
50) Toluene-d8	10.09	98	1112372	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
62) 4-Bromofluorobenzene	12.40	95	361335	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
Target Compounds						
4) Vinyl Chloride	2.19	62	45033	5.876	ug/l	100
27) cis-1,2-Dichloroethene	6.82	96	172756	25.057	ug/l	97
44) Trichloroethene	8.83	130	21891	3.233	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN032219\
Data File : VN054593.D
Acq On : 23 Mar 2019 8:18
Operator : JC/SP
Sample : K2045-19
Misc : 5.93µL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP022SB444-190320-(23-25)

Quant Time: Mar 25 02:25:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032019W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 22 02:09:07 2019
Response via : Initial Calibration

