

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071462.D
 Acq On : 22 Mar 2022 12:55
 Operator : JC\MD
 Sample : PB143494TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143494TB

Quant Time: Mar 23 07:07:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	678730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1168933	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1119181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	396637	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	453312	48.731	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.460%
35) Dibromofluoromethane	8.022	113	341304	46.270	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.540%
50) Toluene-d8	10.433	98	1461355	49.205	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.400%
62) 4-Bromofluorobenzene	12.727	95	534755	53.173	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.340%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	40284	8.900	ug/l	97
18) Methyl Acetate	4.857	43	14990	1.692	ug/l	96
20) Methylene Chloride	5.098	84	11960	1.817	ug/l	93
25) 2-Butanone	7.340	43	9533	1.307	ug/l	97
43) Isopropyl Acetate	8.557	43	91587	3.765	ug/l #	86
95) Naphthalene	15.498	128	13552	5.737	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071462.D
Acq On : 22 Mar 2022 12:55
Operator : JC\MD
Sample : PB143494TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143494TB

Quant Time: Mar 23 07:07:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

