

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071464.D
 Acq On : 22 Mar 2022 13:43
 Operator : JC\MD
 Sample : PB143494ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143494ZHE#02

Quant Time: Mar 23 07:07:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	729338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1263393	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1217351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	455823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	484602	48.480	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.960%
35) Dibromofluoromethane	8.022	113	367862	46.142	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.280%
50) Toluene-d8	10.439	98	1591057	49.566	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.140%
62) 4-Bromofluorobenzene	12.727	95	588748	54.164	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.320%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	39263	8.073	ug/l	96
18) Methyl Acetate	4.857	43	19715	1.965	ug/l #	87
20) Methylene Chloride	5.093	84	11833	1.703	ug/l	95
25) 2-Butanone	7.345	43	7747	0.989	ug/l #	78
43) Isopropyl Acetate	8.557	43	140174	5.332	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071464.D
Acq On : 22 Mar 2022 13:43
Operator : JC\MD
Sample : PB143494ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143494ZHE#02

Quant Time: Mar 23 07:07:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

