

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071470.D
 Acq On : 22 Mar 2022 16:32
 Operator : JC\MD
 Sample : N2061-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-20220321

Quant Time: Mar 23 07:09:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	739543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1278998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1205319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	419302	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	486872	48.035	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.080%		
35) Dibromofluoromethane	8.022	113	372172	46.113	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.220%		
50) Toluene-d8	10.439	98	1571822	48.370	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.740%		
62) 4-Bromofluorobenzene	12.727	95	569668	51.769	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	103.540%		
Target Compounds						
16) Acetone	4.298	43	13092	2.655	ug/l	97
18) Methyl Acetate	4.851	43	13603	1.487	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071470.D
Acq On : 22 Mar 2022 16:32
Operator : JC\MD
Sample : N2061-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-20220321

Quant Time: Mar 23 07:09:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

