

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
 Data File : VN071477.D
 Acq On : 22 Mar 2022 19:22
 Operator : JC\MD
 Sample : N2053-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1271

Quant Time: Mar 23 07:10:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	694246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1202364	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1139883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	414467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	468165	49.203	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.400%
35) Dibromofluoromethane	8.016	113	353799	46.630	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.260%
50) Toluene-d8	10.439	98	1478019	48.382	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.760%
62) 4-Bromofluorobenzene	12.727	95	553133	53.471	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.940%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.363	59	6070	3.247	ug/l #	74
13) Acrolein	4.040	56	12778	6.459	ug/l	97
16) Acetone	4.281	43	307199	66.355	ug/l	97
18) Methyl Acetate	4.857	43	28789	2.764	ug/l	96
25) 2-Butanone	7.334	43	176565	23.675	ug/l	97
69) o-Xylene	12.280	106	12161	0.766	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032222\
Data File : VN071477.D
Acq On : 22 Mar 2022 19:22
Operator : JC\MD
Sample : N2053-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1271

Quant Time: Mar 23 07:10:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

